

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
 Data File : VN074609.D
 Acq On : 21 Sep 2022 13:24
 Operator : JC\MD
 Sample : N4645-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-10

Quant Time: Sep 22 06:57:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	383796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	754195	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	768607	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	355807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	320587	49.230	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.460%
35) Dibromofluoromethane	7.951	113	255709	51.040	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.080%
50) Toluene-d8	10.381	98	951629	49.115	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.240%
62) 4-Bromofluorobenzene	12.674	95	355932	53.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.280%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.405	62	108156	14.385	ug/l #	85
19) Methyl tert-butyl Ether	5.534	73	51815	2.432	ug/l #	73
27) cis-1,2-Dichloroethene	7.251	96	49953	7.029	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
Data File : VN074609.D
Acq On : 21 Sep 2022 13:24
Operator : JC\MD
Sample : N4645-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-10

Quant Time: Sep 22 06:57:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

