

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
 Data File : VN074613.D
 Acq On : 21 Sep 2022 15:01
 Operator : JC\MD
 Sample : PB147776ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147776ZHE#02

Quant Time: Sep 22 06:58:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	379726	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	754559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	819125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	348067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	310756	48.232	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.460%
35) Dibromofluoromethane	7.951	113	258087	51.490	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.980%
50) Toluene-d8	10.380	98	958851	49.464	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.920%
62) 4-Bromofluorobenzene	12.674	95	386984	58.288	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.580%
Target Compounds						
					Qvalue	
16) Acetone	4.216	43	56465	13.696	ug/l	95
18) Methyl Acetate	4.781	43	20842	1.628	ug/l #	79
20) Methylene Chloride	5.022	84	21313	3.415	ug/l	89
37) Ethyl Acetate	7.340	43	103798	7.576	ug/l #	95
43) Isopropyl Acetate	8.492	43	490571	22.858	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
Data File : VN074613.D
Acq On : 21 Sep 2022 15:01
Operator : JC\MD
Sample : PB147776ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147776ZHE#02

Quant Time: Sep 22 06:58:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

