

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
 Data File : VN074616.D
 Acq On : 21 Sep 2022 16:15
 Operator : JC\MD
 Sample : N4566-03DL 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1DL

Quant Time: Sep 22 06:58:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	351261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	689189	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	724204	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	343865	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	294759	49.456	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.920%		
35) Dibromofluoromethane	7.951	113	241504	52.751	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	105.500%		
50) Toluene-d8	10.380	98	864774	48.842	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.680%		
62) 4-Bromofluorobenzene	12.674	95	370084	61.029	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	122.060%		
Target Compounds						Qvalue
31) Cyclohexane	8.022	56	42300	3.999	ug/l #	91
39) Methylcyclohexane	9.392	83	21940	2.277	ug/l	88
52) Toluene	10.445	92	46478	3.163	ug/l	85
67) Ethyl Benzene	11.786	91	635948	19.052	ug/l	100
68) m/p-Xylenes	11.892	106	1526712	123.706	ug/l	96
69) o-Xylene	12.221	106	632243	50.245	ug/l	96
73) Isopropylbenzene	12.521	105	17476	0.496	ug/l	99
78) n-propylbenzene	12.863	91	60676	1.447	ug/l	92
80) 1,3,5-Trimethylbenzene	13.004	105	128596	4.480	ug/l	100
84) 1,2,4-Trimethylbenzene	13.310	105	477527	16.533	ug/l	99
95) Naphthalene	15.439	128	94748	4.706	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092122\
Data File : VN074616.D
Acq On : 21 Sep 2022 16:15
Operator : JC\MD
Sample : N4566-03DL 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1DL

Quant Time: Sep 22 06:58:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

