

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092123\
 Data File : VN079315.D
 Acq On : 21 Sep 2023 09:52
 Operator : JC\MD
 Sample : VN0921MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0921MBL01

Quant Time: Sep 22 02:26:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	179469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	329104	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	289171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	82791	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	126499	52.215	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.420%
35) Dibromofluoromethane	8.171	113	99853	45.613	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.220%
50) Toluene-d8	10.570	98	328387	40.853	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	81.700%#
62) 4-Bromofluorobenzene	12.853	95	103685	40.438	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.880%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092123\
Data File : VN079315.D
Acq On : 21 Sep 2023 09:52
Operator : JC\MD
Sample : VN0921MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0921MBL01

Quant Time: Sep 22 02:26:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

