

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092123\
 Data File : VN079316.D
 Acq On : 21 Sep 2023 10:52
 Operator : JC\MD
 Sample : VN0921WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0921WBL01

Quant Time: Sep 22 02:26:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 12 05:05:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	156742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	293114	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	256862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	70541	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	131058	61.941	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	123.880%
35) Dibromofluoromethane	8.171	113	102095	52.363	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.720%
50) Toluene-d8	10.571	98	335472	46.859	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.720%
62) 4-Bromofluorobenzene	12.853	95	99680	43.650	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092123\
Data File : VN079316.D
Acq On : 21 Sep 2023 10:52
Operator : JC\MD
Sample : VN0921WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0921WBL01

Quant Time: Sep 22 02:26:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091123W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 12 05:05:24 2023
Response via : Initial Calibration

