

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092220\
 Data File : VN063527.D
 Acq On : 22 Sep 2020 12:46
 Operator : JC/MD
 Sample : VN0922MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0922MBL01

Quant Time: Sep 23 04:09:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	219999	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	409996	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	398560	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	166138	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	178597	47.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.38%	
35) Dibromofluoromethane	7.55	113	130948	46.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.28%	
50) Toluene-d8	10.06	98	519384	48.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.34%	
62) 4-Bromofluorobenzene	12.38	95	203225	51.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092220\
Data File : VN063527.D
Acq On : 22 Sep 2020 12:46
Operator : JC/MD
Sample : VN0922MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0922MBL01

Quant Time: Sep 23 04:09:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

