

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063532.D
 Acq On : 22 Sep 2020 14:56
 Operator : JC/MD
 Sample : L4038-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 23 04:21:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	226424	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	423841	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	425745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	182365	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	186576	48.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.82%	
35) Dibromofluoromethane	7.55	113	137311	46.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.60%	
50) Toluene-d8	10.06	98	539967	48.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.88%	
62) 4-Bromofluorobenzene	12.38	95	215951	53.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063532.D
Acq On : 22 Sep 2020 14:56
Operator : JC/MD
Sample : L4038-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 23 04:21:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

