

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063538.D
 Acq On : 22 Sep 2020 17:20
 Operator : JC/MD
 Sample : L4082-19DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE131D3-20200916DL

Quant Time: Sep 23 04:43:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	206627	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	381358	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	388163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	176085	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	173947	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
35) Dibromofluoromethane	7.55	113	126856	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
50) Toluene-d8	10.06	98	493936	49.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.52%	
62) 4-Bromofluorobenzene	12.38	95	204154	56.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.20%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.72	101	152285	66.150	ug/l	98
44) Trichloroethene	8.80	130	13711	4.874	ug/l	99
64) Tetrachloroethene	10.60	164	4338	1.641	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063538.D
Acq On : 22 Sep 2020 17:20
Operator : JC/MD
Sample : L4082-19DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE131D3-20200916DL

Quant Time: Sep 23 04:43:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

