

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063543.D
 Acq On : 22 Sep 2020 19:19
 Operator : JC/MD
 Sample : L4082-14DL 4X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP01-20200915DL

Manual Integrations
 APPROVED

MMDadoda
 9/23/2020 1:05:51 PM

Quant Time: Sep 23 11:49:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	226568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	424583	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	427494	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	178304	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	189005	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
35) Dibromofluoromethane	7.55	113	136176	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
50) Toluene-d8	10.06	98	540660	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	12.38	95	215416	53.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.34%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	2651m	1.050	ug/l	
12) 1,1-Dichloroethene	3.70	96	1362	0.533	ug/l #	55
30) Chloroform	7.33	83	1363	0.241	ug/l #	83
44) Trichloroethene	8.80	130	112952	36.065	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063543.D
Acq On : 22 Sep 2020 19:19
Operator : JC/MD
Sample : L4082-14DL 4X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP01-20200915DL

Manual Integrations
APPROVED

MMDadoda
9/23/2020 1:05:51 PM

Quant Time: Sep 23 11:49:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

