

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063546.D
 Acq On : 22 Sep 2020 20:31
 Operator : JC/MD
 Sample : L4082-13DL 40X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE108D2-20200915DL

Quant Time: Sep 23 05:13:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	220430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	413773	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	416092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	177102	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	185738	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
35) Dibromofluoromethane	7.55	113	133103	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
50) Toluene-d8	10.06	98	528691	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
62) 4-Bromofluorobenzene	12.38	95	208381	52.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.56%	
Target Compounds						
44) Trichloroethene	8.80	130	196793	64.477	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063546.D
Acq On : 22 Sep 2020 20:31
Operator : JC/MD
Sample : L4082-13DL 40X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE108D2-20200915DL

Quant Time: Sep 23 05:13:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

