

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063549.D
 Acq On : 22 Sep 2020 21:42
 Operator : JC/MD
 Sample : L4094-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KTS-WC-L-01

Manual Integrations
 APPROVED

MMDadoda
 9/23/2020 1:06:31 PM

Quant Time: Sep 23 05:25:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	217638	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	406119	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	407220	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	171117	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	182628	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
35) Dibromofluoromethane	7.55	113	134376	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
50) Toluene-d8	10.06	98	517577	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	12.38	95	203177	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
Target Compounds						
16) Acetone	3.78	43	39046	32.794	ug/l	98
25) 2-Butanone	6.80	43	10379m	5.668	ug/l	
44) Trichloroethene	8.81	130	2269m	0.757	ug/l	
52) Toluene	10.13	92	2591	0.349	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063549.D
Acq On : 22 Sep 2020 21:42
Operator : JC/MD
Sample : L4094-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KTS-WC-L-01

Manual Integrations
APPROVED

MMDadoda
9/23/2020 1:06:31 PM

Quant Time: Sep 23 05:25:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

