

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063559.D
 Acq On : 23 Sep 2020 2:51
 Operator : JC/MD
 Sample : PB131779ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131779ZHE#11

Quant Time: Sep 23 07:12:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	215882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	409209	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	407181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	175609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	184247	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
35) Dibromofluoromethane	7.55	113	132430	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
50) Toluene-d8	10.06	98	519280	48.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.50%	
62) 4-Bromofluorobenzene	12.38	95	206001	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
Target Compounds						
20) Methylene Chloride	4.51	84	4326	1.446	ug/l #	Qvalue 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063559.D
Acq On : 23 Sep 2020 2:51
Operator : JC/MD
Sample : PB131779ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131779ZHE#11

Quant Time: Sep 23 07:12:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

