

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063560.D
 Acq On : 23 Sep 2020 3:15
 Operator : JC/MD
 Sample : PB131779ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131779ZHE#12

Manual Integrations
 APPROVED

MMDadoda
 9/23/2020 1:06:42 PM

Quant Time: Sep 23 07:13:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	204531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	389297	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	387138	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	161633	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	173150	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
35) Dibromofluoromethane	7.55	113	127330	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
50) Toluene-d8	10.06	98	494330	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
62) 4-Bromofluorobenzene	12.38	95	186804	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
Target Compounds						
20) Methylene Chloride	4.51	84	4208m	1.484	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063560.D
Acq On : 23 Sep 2020 3:15
Operator : JC/MD
Sample : PB131779ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131779ZHE#12

Manual Integrations
APPROVED

MMDadoda
9/23/2020 1:06:42 PM

Quant Time: Sep 23 07:13:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

