

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063565.D
 Acq On : 23 Sep 2020 5:16
 Operator : JC/MD
 Sample : PB131772ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131772ZHE#03

Quant Time: Sep 23 07:21:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	208982	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	401828	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	414705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	172167	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	179212	50.38	ug/l	0.00
Spiked Amount				50.000		
Recovery				=		100.76%
35) Dibromofluoromethane	7.55	113	129515	46.56	ug/l	0.00
Spiked Amount				50.000		
Recovery				=		93.12%
50) Toluene-d8	10.06	98	524449	50.14	ug/l	0.00
Spiked Amount				50.000		
Recovery				=		100.28%
62) 4-Bromofluorobenzene	12.38	95	206283	53.80	ug/l	0.00
Spiked Amount				50.000		
Recovery				=		107.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	13348	11.675	ug/l	100
18) Methyl Acetate	4.28	43	3023	1.042	ug/l	96
20) Methylene Chloride	4.51	84	7681	2.652	ug/l #	88
43) Isopropyl Acetate	8.13	43	84607	11.262	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063565.D
 Acq On : 23 Sep 2020 5:16
 Operator : JC/MD
 Sample : PB131772ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131772ZHE#03

Quant Time: Sep 23 07:21:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

