

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\  
 Data File : VN063568.D  
 Acq On : 23 Sep 2020 6:28  
 Operator : JC/MD  
 Sample : PB131772ZHE#06  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 47 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB131772ZHE#06

Quant Time: Sep 23 07:34:32 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N091620W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Sep 16 13:05:20 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 7.62  | 168  | 209358   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.55  | 114  | 415714   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 11.38 | 117  | 414639   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.32 | 152  | 167253   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 7.99  | 65   | 181638   | 50.97  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 101.94% |          |
| 35) Dibromofluoromethane    | 7.55  | 113  | 130061   | 45.19  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 90.38%  |          |
| 50) Toluene-d8              | 10.06 | 98   | 525386   | 48.55  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 97.10%  |          |
| 62) 4-Bromofluorobenzene    | 12.38 | 95   | 204023   | 51.43  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 102.86% |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 3.78  | 43   | 34993    | 30.552 | ug/l    | 97       |
| 18) Methyl Acetate          | 4.29  | 43   | 4127     | 1.419  | ug/l    | 97       |
| 20) Methylene Chloride      | 4.51  | 84   | 18911    | 6.517  | ug/l    | 88       |
| 43) Isopropyl Acetate       | 8.13  | 43   | 77663    | 9.992  | ug/l    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\  
Data File : VN063568.D  
Acq On : 23 Sep 2020 6:28  
Operator : JC/MD  
Sample : PB131772ZHE#06  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 47 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
PB131772ZHE#06

Quant Time: Sep 23 07:34:32 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_N\METHODS\82N091620W.M  
Quant Title : SW846 8260  
QLast Update : Wed Sep 16 13:05:20 2020  
Response via : Initial Calibration

