

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063569.D
 Acq On : 23 Sep 2020 6:52
 Operator : JC/MD
 Sample : PB131772ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131772ZHE#07

Manual Integrations
 APPROVED

MMDadoda
 9/23/2020 1:06:52 PM

Quant Time: Sep 23 07:38:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	205580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	396841	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	399467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	168514	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	178382	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
35) Dibromofluoromethane	7.55	113	126905	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
50) Toluene-d8	10.06	98	515660	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
62) 4-Bromofluorobenzene	12.38	95	200072	52.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.66%	
Target Compounds						
16) Acetone	3.78	43	33697	29.961	ug/l	97
18) Methyl Acetate	4.29	43	4711m	1.650	ug/l	
20) Methylene Chloride	4.50	84	18670	6.552	ug/l	96
43) Isopropyl Acetate	8.13	43	77277	10.415	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063569.D
Acq On : 23 Sep 2020 6:52
Operator : JC/MD
Sample : PB131772ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131772ZHE#07

Manual Integrations
APPROVED

MMDadoda
9/23/2020 1:06:52 PM

Quant Time: Sep 23 07:38:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

