

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092220\
 Data File : VN063572.D
 Acq On : 23 Sep 2020 8:04
 Operator : JC/MD
 Sample : L3872-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BU-02-092120

Manual Integrations
 APPROVED

MMDadoda
 9/23/2020 1:06:56 PM

Quant Time: Sep 23 08:44:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	204239	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	395194	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	399094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	164018	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	177761	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
35) Dibromofluoromethane	7.55	113	126743	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
50) Toluene-d8	10.06	98	508403	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	12.38	95	198595	52.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.32%	
Target Compounds						
16) Acetone	3.78	43	16231	14.526	ug/l	98
18) Methyl Acetate	4.28	43	3540m	1.248	ug/l	
20) Methylene Chloride	4.51	84	11816	4.174	ug/l	97
43) Isopropyl Acetate	8.13	43	74915	10.139	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092220\
Data File : VN063572.D
Acq On : 23 Sep 2020 8:04
Operator : JC/MD
Sample : L3872-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BU-02-092120

Manual Integrations
APPROVED

MMDadoda
9/23/2020 1:06:56 PM

Quant Time: Sep 23 08:44:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

