

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092220\
 Data File : VN063551.D
 Acq On : 22 Sep 2020 22:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 23 05:31:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.566	0.508	10.2	84	0.00
3 P	Chloromethane	0.749	0.614	18.0	80	0.00
4 C	Vinyl Chloride	0.664	0.628	5.4#	87	0.00
5 T	Bromomethane	0.394	0.376	4.6	93	0.00
6 T	Chloroethane	0.408	0.376	7.8	91	0.00
7 T	Trichlorofluoromethane	0.991	0.977	1.4	92	0.00
8 T	Diethyl Ether	0.386	0.358	7.3	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.523	6.1	88	0.00
10 T	Methyl Iodide	0.725	0.678	6.5	86	0.00
11 T	Tert butyl alcohol	0.118	0.093	21.2#	74	0.00
12 CM	1,1-Dichloroethene	0.564	0.514	8.9#	85	0.00
13 T	Acrolein	0.044	0.061	-38.6#	125	0.00
14 T	Allyl chloride	1.116	0.881	21.1#	72	0.00
15 T	Acrylonitrile	0.300	0.278	7.3	85	0.00
16 T	Acetone	0.274	0.246	10.2	88	0.00
17 T	Carbon Disulfide	1.641	1.373	16.3	78	0.00
18 T	Methyl Acetate	0.694	0.666	4.0	91	0.00
19 T	Methyl tert-butyl Ether	2.103	1.885	10.4	85	0.00
20 T	Methylene Chloride	0.693	0.625	9.8	89	0.00
21 T	trans-1,2-Dichloroethene	0.625	0.572	8.5	87	0.00
22 T	Diisopropyl ether	2.276	2.085	8.4	87	0.00
23 T	Vinyl Acetate	1.869	1.718	8.1	83	0.00
24 P	1,1-Dichloroethane	1.243	1.143	8.0	86	0.00
25 T	2-Butanone	0.421	0.376	10.7	83	0.00
26 T	2,2-Dichloropropane	1.094	0.853	22.0#	75	0.00
27 T	cis-1,2-Dichloroethene	0.708	0.671	5.2	87	0.00
28 T	Bromochloromethane	0.574	0.530	7.7	90	0.00
29 T	Tetrahydrofuran	0.287	0.251	12.5	84	0.00
30 C	Chloroform	1.247	1.197	4.0#	90	0.00
31 T	Cyclohexane	1.176	0.950	19.2	80	0.00
32 T	1,1,1-Trichloroethane	1.109	1.041	6.1	89	0.00
33 S	1,2-Dichloroethane-d4	0.851	0.762	10.5	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.346	0.312	9.8	88	0.00
36 T	1,1-Dichloropropene	0.533	0.484	9.2	85	0.00
37 T	Ethyl Acetate	0.523	0.484	7.5	85	0.00
38 T	Carbon Tetrachloride	0.544	0.518	4.8	88	0.00
39 T	Methylcyclohexane	0.578	0.489	15.4	79	0.00
40 TM	Benzene	1.521	1.431	5.9	88	0.00
41 T	Methacrylonitrile	0.235	0.254	-8.1	99	0.00
42 TM	1,2-Dichloroethane	0.612	0.580	5.2	88	0.00
43 T	Isopropyl Acetate	0.935	0.822	12.1	82	0.00
44 TM	Trichloroethene	0.369	0.350	5.1	90	0.00
45 C	1,2-Dichloropropane	0.414	0.388	6.3#	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092220\
 Data File : VN063551.D
 Acq On : 22 Sep 2020 22:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 23 05:31:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.272	0.268	1.5	91	0.00
47 T	Bromodichloromethane	0.565	0.556	1.6	91	0.00
48 T	Methyl methacrylate	0.459	0.406	11.5	82	0.00
49 T	1,4-Dioxane	0.007	0.005	28.6#	78	0.00
50 S	Toluene-d8	1.302	1.170	10.1	86	0.00
51 T	4-Methyl-2-Pentanone	0.523	0.489	6.5	86	0.00
52 CM	Toluene	0.914	0.889	2.7#	88	0.00
53 T	t-1,3-Dichloropropene	0.631	0.579	8.2	84	0.00
54 T	cis-1,3-Dichloropropene	0.666	0.609	8.6	83	0.00
55 T	1,1,2-Trichloroethane	0.363	0.360	0.8	91	0.00
56 T	Ethyl methacrylate	0.567	0.540	4.8	87	0.00
57 T	1,3-Dichloropropane	0.655	0.613	6.4	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.278	6.4	89	0.00
59 T	2-Hexanone	0.379	0.355	6.3	86	0.00
60 T	Dibromochloromethane	0.410	0.393	4.1	88	0.00
61 T	1,2-Dibromoethane	0.377	0.372	1.3	90	0.00
62 S	4-Bromofluorobenzene	0.477	0.454	4.8	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.340	0.342	-0.6	98	0.00
65 PM	Chlorobenzene	1.061	1.009	4.9	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.381	2.1	91	0.00
67 C	Ethyl Benzene	1.999	1.901	4.9#	90	0.00
68 T	m/p-Xylenes	0.721	0.689	4.4	90	0.00
69 T	o-Xylene	0.686	0.663	3.4	90	0.00
70 T	Styrene	1.151	1.148	0.3	90	0.00
71 P	Bromoform	0.290	0.276	4.8	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	4.007	3.810	4.9	89	0.00
74 T	N-amyl acetate	1.897	1.674	11.8	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.150	7.3	89	0.00
76 T	1,2,3-Trichloropropane	1.238	1.233	0.4	101	0.00
77 T	Bromobenzene	0.901	0.898	0.3	93	0.00
78 T	n-propylbenzene	4.667	4.451	4.6	89	0.00
79 T	2-Chlorotoluene	2.797	2.680	4.2	90	0.00
80 T	1,3,5-Trimethylbenzene	3.341	3.260	2.4	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.451	0.380	15.7	81	0.00
82 T	4-Chlorotoluene	2.956	2.830	4.3	91	0.00
83 T	tert-Butylbenzene	2.758	2.591	6.1	88	0.00
84 T	1,2,4-Trimethylbenzene	3.375	3.258	3.5	88	0.00
85 T	sec-Butylbenzene	3.691	3.441	6.8	87	0.00
86 T	p-Isopropyltoluene	3.256	3.089	5.1	87	0.00
87 T	1,3-Dichlorobenzene	1.671	1.615	3.4	90	0.00
88 T	1,4-Dichlorobenzene	1.698	1.616	4.8	88	0.00
89 T	n-Butylbenzene	3.001	2.765	7.9	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092220\
Data File : VN063551.D
Acq On : 22 Sep 2020 22:30
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 23 05:31:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.620	0.560	9.7	84	0.00
91 T	1,2-Dichlorobenzene	1.640	1.593	2.9	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.258	12.5	80	0.00
93 T	1,2,4-Trichlorobenzene	0.967	0.887	8.3	82	0.00
94 T	Hexachlorobutadiene	0.400	0.401	-0.3	92	0.00
95 T	Naphthalene	3.194	2.863	10.4	80	0.00
96 T	1,2,3-Trichlorobenzene	0.934	0.849	9.1	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6