


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092220\
 Data File : VN063552.D
 Acq On : 22 Sep 2020 23:17
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Title : SW846 8260
 Last Update : Wed Sep 16 13:05:20 2020



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3294

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.1	12233	PASS
75	95	30	60	56.4	29872	PASS
95	95	100	100	100.0	52970	PASS
96	95	5	9	7.4	3915	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.5	38954	PASS
175	174	5	9	8.6	3348	PASS
176	174	95	101	96.0	37389	PASS
177	176	5	9	6.5	2412	PASS