

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092225\
 Data File : VN087915.D
 Acq On : 22 Sep 2025 12:26
 Operator : JC\MD
 Sample : VN0922WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0922WBL01

Quant Time: Sep 23 04:51:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090425W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Sep 05 03:29:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	53463	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	267356	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.841	117	245968	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	131878	28.876	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.267%
60) 4-Bromofluorobenzene	12.829	95	100591	25.434	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	84.767%
63) Toluene-d8	10.547	98	329123	28.772	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	95.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN092225\
Data File : VN087915.D
Acq On : 22 Sep 2025 12:26
Operator : JC\MD
Sample : VN0922WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0922WBL01

Quant Time: Sep 23 04:51:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090425W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Sep 05 03:29:53 2025
Response via : Initial Calibration

