

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051429.D
 Acq On : 23 Sep 2018 22:33
 Operator : MD\SY
 Sample : PB113124TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124TB

Quant Time: Sep 24 06:49:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	629973	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	988068	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	822357	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	281842	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	438426	55.56	ug/l	0.00
Spiked Amount			Recovery	=	111.12%	
35) Dibromofluoromethane	7.59	113	386017	48.86	ug/l	0.00
Spiked Amount			Recovery	=	97.72%	
50) Toluene-d8	10.09	98	1360825	45.74	ug/l	0.00
Spiked Amount			Recovery	=	91.48%	
62) 4-Bromofluorobenzene	12.40	95	360984	37.50	ug/l	0.00
Spiked Amount			Recovery	=	75.00%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.57	94	2187	0.412	ug/l	83
8) Diethyl Ether	3.41	74	107	Below Cal	#	1
10) Methyl Iodide	3.94	142	463	0.322	ug/l #	58
11) Tert butyl alcohol	4.79	59	96	0.216	ug/l #	78
16) Acetone	3.83	43	23296	Below Cal		99
17) Carbon Disulfide	4.06	76	6346	0.308	ug/l	97
18) Methyl Acetate	4.34	43	12169	2.025	ug/l #	84
20) Methylene Chloride	4.55	84	14101	1.806	ug/l	99
25) 2-Butanone	6.84	43	2392	0.886	ug/l	95
29) Tetrahydrofuran	7.21	42	635	0.371	ug/l #	38
31) Cyclohexane	7.67	56	11537	0.928	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	44011	3.846	ug/l #	47
38) Carbon Tetrachloride	7.67	117	57172	4.777	ug/l #	17
43) Isopropyl Acetate	8.17	43	239019	20.092	ug/l	97
48) Methyl methacrylate	9.18	41	5704	0.962	ug/l #	21
51) 4-Methyl-2-Pentanone	9.98	43	1506	0.230	ug/l #	81
58) 2-Chloroethyl Vinyl ether	9.71	63	213	20.409	ug/l #	48
59) 2-Hexanone	10.77	43	16548	3.902	ug/l #	57
71) Bromoform	12.41	173	1975	0.362	ug/l #	1
76) 1,2,3-Trichloropropane	12.40	75	179044	38.267	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	179044	117.670	ug/l #	12
93) 1,2,4-Trichlorobenzene	14.91	180	2578	6.398	ug/l	88
94) Hexachlorobutadiene	15.01	225	1184	0.310	ug/l	89
95) Naphthalene	15.13	128	5099	8.453	ug/l	96
96) 1,2,3-Trichlorobenzene	15.31	180	2873	0.592	ug/l	88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051429.D
Acq On : 23 Sep 2018 22:33
Operator : MD\SY
Sample : PB113124TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113124TB

Quant Time: Sep 24 06:49:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampleId :


PB113124TB

Tgt Ion:168 Resp: 629973

Ion Ratio Lower Upper

168 100

99 54.1 40.2 60.2



Abundance Ion 167.90 (167.60 to 168.60): VN051393.D (-1813) (-)

Ion 98.90 (98.60 to 99.60): VN051393.D (-1813) (-)

7.67

300000

250000

200000



150000

100000

50000

0

Time-->



#5

Bromomethane

Concen: 0.412 ug/l

RT: 2.57 min Scan# 249

Delta R.T. 0.01 min

Lab File: VN051429.D


Acq: 23 Sep 2018 22:33

Tgt Ion: 94 Resp: 2187

Ion Ratio Lower Upper

94 100

96 75.9 73.6 110.4



Abundance Ion 93.90 (93.60 to 94.60): VN051393.D (-228) (-)

Ion 95.90 (95.60 to 96.60): VN051393.D (-228) (-)

2.57

1200

1000

800


600


400

200

0

Time-->





#8

Diethyl Ether

Concen: Below Cal

RT: 3.41 min Scan# 508

Delta R.T. -0.01 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampleId :





PB113124TB

Tgt Ion: 74 Resp: 107

Ion Ratio Lower Upper

74 100

45 386.9 50.2 150.7#



#10

Methyl Iodide

Concen: 0.322 ug/l

RT: 3.94 min Scan# 675

Delta R.T. -0.01 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33




Tgt Ion: 142 Resp: 463


Ion Ratio Lower Upper

142 100

127 13.4 33.2 49.8#

141 0.0 11.5 17.3#









#11
Tert butyl alcohol
Concen: 0.216 ug/l
RT: 4.79 min Scan# 938
Delta R.T. 0.00 min
Lab File: VN051429.D
Acq: 23 Sep 2018 22:33




Instrument :
MSVOA_N
ClientSampleId :
PB113124TB


Tgt Ion: 59 Resp: 96
Ion Ratio Lower Upper
59 100
57 0.0 6.2 9.2#



#16
Acetone
Concen: Below Cal
RT: 3.83 min Scan# 639
Delta R.T. 0.01 min
Lab File: VN051429.D
Acq: 23 Sep 2018 22:33

Tgt Ion: 43 Resp: 23296
Ion Ratio Lower Upper
43 100
58 31.8 25.0 37.4









#17
Carbon Disulfide
Concen: 0.308 ug/l
RT: 4.06 min Scan# 710
Delta R.T. 0.00 min
Lab File: VN051429.D
Acq: 23 Sep 2018 22:33




Instrument :
MSVOA_N
ClientSampleId :
PB113124TB


Tgt Ion: 76 Resp: 6346
Ion Ratio Lower Upper
76 100
78 10.2 7.3 10.9



#18
Methyl Acetate
Concen: 2.025 ug/l
RT: 4.34 min Scan# 797
Delta R.T. 0.01 min
Lab File: VN051429.D
Acq: 23 Sep 2018 22:33

Tgt Ion: 43 Resp: 12169
Ion Ratio Lower Upper
43 100
74 15.7 18.8 28.2#





#20

Methylene Chloride

Concen: 1.806 ug/l

RT: 4.55 min Scan# 865

Delta R.T. 0.00 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampled :

PB113124TB

Tgt Ion: 84 Resp: 14101


Ion Ratio Lower Upper

84 100

49 125.0 99.9 149.9

51 36.6 31.4 47.2

86 63.2 50.9 76.3






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#25

2-Butanone

Concen: 0.886 ug/l

RT: 6.84 min Scan# 1576

Delta R.T. 0.00 min

Lab File: VN051429.D


Acq: 23 Sep 2018 22:33

Tgt Ion: 43 Resp: 2392

Ion Ratio Lower Upper



43 100


72 29.9 21.8 32.6



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0









#29
Tetrahydrofuran
Concen: 0.371 ug/l
RT: 7.21 min Scan# 1692
Delta R.T. 0.00 min
Lab File: VN051429.D
Acq: 23 Sep 2018 22:33




Instrument :
MSVOA_N
ClientSampleId :
PB113124TB


Tgt Ion: 42 Resp: 635
Ion Ratio Lower Upper
42 100
72 5.0 34.3 51.5#
71 0.0 32.2 48.4#



#31
Cyclohexane
Concen: 0.928 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. 0.01 min
Lab File: VN051429.D
Acq: 23 Sep 2018 22:33

Tgt Ion: 56 Resp: 11537
Ion Ratio Lower Upper
56 100
69 206.2 25.8 38.6#
84 149.9 66.0 99.0#









#33
 1,2-Dichloroethane-d4
 Concen: 55.556 ug/l
 RT: 8.03 min Scan# 1946
 Delta R.T. 0.00 min
 Lab File: VN051429.D
 Acq: 23 Sep 2018 22:33




Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124TB


Tgt Ion: 65 Resp: 438426
 Ion Ratio Lower Upper
 65 100
 67 52.9 0.0 107.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN051429.D
 Acq: 23 Sep 2018 22:33

Tgt Ion: 114 Resp: 988068
 Ion Ratio Lower Upper
 114 100
 63 20.1 0.0 40.2
 88 15.9 0.0 31.4









#35
 Dibromofluoromethane
 Concen: 48.856 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN051429.D
 Acq: 23 Sep 2018 22:33




Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124TB


Tgt Ion	Ratio	Lower	Upper
113	100		
111	101.2	82.6	123.8
192	22.6	18.0	27.0



#36
 1,1-Dichloropropene
 Concen: 3.846 ug/l
 RT: 7.67 min Scan# 1834
 Delta R.T. -0.12 min
 Lab File: VN051429.D
 Acq: 23 Sep 2018 22:33

Tgt Ion	Ratio	Lower	Upper
75	100		
110	7.6	18.1	54.3#
77	0.0	25.6	38.4#





#38

Carbon Tetrachloride

Concen: 4.777 ug/l

RT: 7.67 min Scan# 1834

Delta R.T. -0.11 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampleId :

PB113124TB


Tgt Ion:117 Resp: 57172

Ion Ratio Lower Upper

117 100

119 6.4 76.9 115.3#




121 0.0 24.6 37.0#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 20.092 ug/l

RT: 8.17 min Scan# 1988

Delta R.T. 0.00 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33


Tgt Ion: 43 Resp: 239019

Ion Ratio Lower Upper

43 100

61 19.1 16.8 25.2



87 12.4 10.7 16.1




Abundance Ion 43.00 (42.70 to 43.70): V

Ion 61.00 (60.70 to 61.70): V

Ion 87.00 (86.70 to 87.70): V





#48

Methyl methacrylate

Concen: 0.962 ug/l

RT: 9.18 min Scan# 2303

Delta R.T. -0.02 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampleId :

PB113124TB


Tgt Ion: 41 Resp: 5704

Ion Ratio Lower Upper

41 100

69 0.0 70.4 105.6#




39 16.4 43.8 65.8#



Abundance Ion 41.00 (40.70 to 41.70): VN051393.D (-2298) (-)

Ion 69.00 (68.70 to 69.70): VN051393.D (-2298) (-)

Ion 39.00 (38.70 to 39.70): VN051393.D (-2298) (-)



#50

Toluene-d8

Concen: 45.738 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN051429.D


Acq: 23 Sep 2018 22:33

Tgt Ion: 98 Resp: 1360825

Ion Ratio Lower Upper



98 100


100 63.9 50.6 75.8



Abundance Ion 98.00 (97.70 to 98.70): VN051393.D (-2571) (-)

Ion 100.00 (99.70 to 100.70): VN051393.D (-2571) (-)









#51
 4-Methyl-2-Pentanone
 Concen: 0.230 ug/l
 RT: 9.98 min Scan# 2553
 Delta R.T. -0.00 min
 Lab File: VN051429.D
 Acq: 23 Sep 2018 22:33




Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124TB


Tgt Ion: 43 Resp: 1506
 Ion Ratio Lower Upper
 43 100
 58 27.5 31.5 47.3#



#58
 2-Chloroethyl Vinyl ether
 Concen: 20.409 ug/l
 RT: 9.71 min Scan# 2469
 Delta R.T. 0.02 min
 Lab File: VN051429.D
 Acq: 23 Sep 2018 22:33

Tgt Ion: 63 Resp: 213
 Ion Ratio Lower Upper
 63 100
 106 0.0 21.4 32.2#





#59

2-Hexanone

Concen: 3.902 ug/l

RT: 10.77 min Scan# 2797

Delta R.T. 0.01 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampleId :


PB113124TB

Tgt Ion: 43 Resp: 16548

Ion Ratio Lower Upper




43 100

58 24.2 27.7 83.0#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#62

4-Bromofluorobenzene

Concen: 37.505 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33


Tgt Ion: 95 Resp: 360984

Ion Ratio Lower Upper

95 100

174 90.5 0.0 181.6



176 88.1 0.0 176.4




Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN051429.D
Acq: 23 Sep 2018 22:33




Instrument :
MSVOA_N
ClientSampleId :
PB113124TB


Tgt Ion:117 Resp: 822357
Ion Ratio Lower Upper
117 100
82 54.8 44.2 66.2
119 31.4 25.9 38.9



#71
Bromoform
Concen: 0.362 ug/l
RT: 12.41 min Scan# 3307
Delta R.T. 0.28 min
Lab File: VN051429.D
Acq: 23 Sep 2018 22:33

Tgt Ion:173 Resp: 1975
Ion Ratio Lower Upper
173 100
175 1094.3 24.8 74.3#
254 0.0 0.2 0.2#





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampleId :

PB113124TB


Tgt Ion:152 Resp: 281842

Ion Ratio Lower Upper

152 100

115 55.1 28.3 84.9




150 156.2 0.0 353.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 38.267 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. -0.15 min

Lab File: VN051429.D


Acq: 23 Sep 2018 22:33

Tgt Ion: 75 Resp: 179044

Ion Ratio Lower Upper



75 100


77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 77.00 (76.70 to 77.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 117.670 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.10 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampleId :

PB113124TB


Tgt Ion: 75 Resp: 179044

Ion Ratio Lower Upper

75 100

53 0.1 73.9 110.9#




89 0.0 36.7 55.1#



Abundance Ion 74.90 (74.60 to 75.60): VN051429.D (-3181) (-)

Ion 53.00 (52.70 to 53.70): VN051429.D (-3181) (-)

Ion 88.90 (88.60 to 89.60): VN051429.D (-3181) (-)



#93

1,2,4-Trichlorobenzene

Concen: 6.398 ug/l

RT: 14.91 min Scan# 4086

Delta R.T. 0.00 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33


Tgt Ion: 180 Resp: 2578

Ion Ratio Lower Upper

180 100

182 82.4 48.5 145.6



145 27.3 14.1 42.4




Abundance Ion 179.80 (179.50 to 180.50): VN051429.D (-3991) (-)

Ion 181.80 (181.50 to 182.50): VN051429.D (-3991) (-)

Ion 144.80 (144.50 to 145.50): VN051429.D (-3991) (-)





#94

Hexachlorobutadiene

Concen: 0.310 ug/l

RT: 15.01 min Scan# 4116

Delta R.T. -0.00 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33

Instrument :

MSVOA_N

ClientSampleId :

PB113124TB


Tgt Ion:225 Resp: 1184

Ion Ratio Lower Upper

225 100

223 53.1 31.9 95.9




227 58.9 32.5 97.5



Abundance Ion 224.70 (224.40 to 225.40): V

Ion 222.70 (222.40 to 223.40): V

Ion 226.70 (226.40 to 227.40): V



#95

Naphthalene

Concen: 8.453 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN051429.D

Acq: 23 Sep 2018 22:33


Tgt Ion:128 Resp: 5099

Ion Ratio Lower Upper

128 100

127 15.3 10.5 15.7



129 9.7 8.6 12.8




Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V





#96
 1,2,3-Trichlorobenzene
 Concen: 0.592 ug/l
 RT: 15.31 min Scan# 4211
 Delta R.T. 0.00 min
 Lab File: VN051429.D
 Acq: 23 Sep 2018 22:33

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124TB

Tgt Ion	Ratio	Lower	Upper
180	100		
182	80.5	47.2	141.6
145	30.6	14.6	43.8

