

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051406.D
 Acq On : 23 Sep 2018 12:20
 Operator : MD\SY
 Sample : J4777-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-02-0920-18-B

Quant Time: Sep 24 07:25:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	689623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1029962	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	882731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	352741	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	451199	52.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.46%	
35) Dibromofluoromethane	7.59	113	401107	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
50) Toluene-d8	10.09	98	1459516	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
62) 4-Bromofluorobenzene	12.40	95	412321	41.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.20%	
Target Compounds						
20) Methylene Chloride	4.55	84	1610769	188.428	ug/l	99
43) Isopropyl Acetate	8.17	43	201426	16.243	ug/l	98
95) Naphthalene	15.13	128	122295	15.869	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051406.D
Acq On : 23 Sep 2018 12:20
Operator : MD\SY
Sample : J4777-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-02-0920-18-B

Quant Time: Sep 24 07:25:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

