

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051409.D
 Acq On : 23 Sep 2018 13:34
 Operator : MD\SY
 Sample : J4777-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-03-092018-A

Quant Time: Sep 24 07:16:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	660810	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	997467	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	829288	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	289843	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	435718	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
35) Dibromofluoromethane	7.59	113	387412	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
50) Toluene-d8	10.09	98	1386125	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
62) 4-Bromofluorobenzene	12.40	95	371850	38.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.54%	
Target Compounds						
20) Methylene Chloride	4.55	84	8667165	1058.095	ug/l	99
30) Chloroform	7.37	83	14227	1.009	ug/l	100
43) Isopropyl Acetate	8.17	43	195303	16.263	ug/l	97
95) Naphthalene	15.13	128	31153	10.470	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051409.D
Acq On : 23 Sep 2018 13:34
Operator : MD\SY
Sample : J4777-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-03-092018-A

Quant Time: Sep 24 07:16:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

