

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051411.D
 Acq On : 23 Sep 2018 14:23
 Operator : MD\SY
 Sample : J4777-24
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-03-092018-C

Quant Time: Sep 24 07:29:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	672879	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1037945	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	864538	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	324490	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	448175	53.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.34%	
35) Dibromofluoromethane	7.59	113	396416	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
50) Toluene-d8	10.09	98	1428938	45.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.44%	
62) 4-Bromofluorobenzene	12.40	95	402815	39.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.68%	
Target Compounds						
18) Methyl Acetate	4.34	43	10442	1.627	ug/l	91
20) Methylene Chloride	4.55	84	554436	66.472	ug/l	99
43) Isopropyl Acetate	8.17	43	226448	18.121	ug/l	95
95) Naphthalene	15.13	128	40605	10.868	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051411.D
Acq On : 23 Sep 2018 14:23
Operator : MD\SY
Sample : J4777-24
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-03-092018-C

Quant Time: Sep 24 07:29:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

