

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051413.D
 Acq On : 23 Sep 2018 15:12
 Operator : MD\SY
 Sample : J4777-28
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-04-092018-B

Quant Time: Sep 24 07:33:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	666517	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1022311	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	853724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	328853	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	445419	53.35	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	106.70%	
35) Dibromofluoromethane	7.59	113	386566	47.29	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	94.58%	
50) Toluene-d8	10.09	98	1413095	45.90	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.80%	
62) 4-Bromofluorobenzene	12.40	95	390964	39.26	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	78.52%	

Target Compounds

						Qvalue
18) Methyl Acetate	4.34	43	11411	1.795	ug/l	98
20) Methylene Chloride	4.55	84	1276686	154.524	ug/l	100
43) Isopropyl Acetate	8.17	43	263999	21.449	ug/l	93
95) Naphthalene	15.13	128	23224	9.638	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051413.D
 Acq On : 23 Sep 2018 15:12
 Operator : MD\SY
 Sample : J4777-28
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-04-092018-B

Quant Time: Sep 24 07:33:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

