

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051414.D
 Acq On : 23 Sep 2018 15:36
 Operator : MD\SY
 Sample : J4777-30
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-04-092018-C

Quant Time: Sep 24 07:35:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	648876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1002539	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	794463	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	268310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	435939	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
35) Dibromofluoromethane	7.59	113	388635	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
50) Toluene-d8	10.09	98	1361550	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
62) 4-Bromofluorobenzene	12.40	95	353863	36.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.46%	
Target Compounds						
18) Methyl Acetate	4.34	43	9916	1.602	ug/l	93
20) Methylene Chloride	4.55	84	1456842	181.123	ug/l	100
43) Isopropyl Acetate	8.17	43	244011	20.216	ug/l	94
95) Naphthalene	15.13	128	9270	8.824	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051414.D
Acq On : 23 Sep 2018 15:36
Operator : MD\SY
Sample : J4777-30
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-04-092018-C

Quant Time: Sep 24 07:35:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

