

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051416.D
 Acq On : 23 Sep 2018 16:26
 Operator : MD\SY
 Sample : J5021-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LS-SOIL-A

Quant Time: Sep 24 07:41:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	656014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1015574	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	851590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	316887	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	443846	54.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.02%	
35) Dibromofluoromethane	7.59	113	392175	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
50) Toluene-d8	10.09	98	1395652	45.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.28%	
62) 4-Bromofluorobenzene	12.40	95	386986	39.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.24%	
Target Compounds						
18) Methyl Acetate	4.33	43	15839	2.531	ug/l #	71
20) Methylene Chloride	4.56	84	55749	6.856	ug/l	95
43) Isopropyl Acetate	8.17	43	245770	20.100	ug/l	94
95) Naphthalene	15.13	128	13455	9.003	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051416.D
Acq On : 23 Sep 2018 16:26
Operator : MD\SY
Sample : J5021-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LS-SOIL-A

Quant Time: Sep 24 07:41:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

