

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051421.D
 Acq On : 23 Sep 2018 18:28
 Operator : MD\SY
 Sample : J5023-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NB-PIT-1

Quant Time: Sep 24 07:49:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	652165	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1016535	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	848393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	317321	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	443554	54.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.58%	
35) Dibromofluoromethane	7.59	113	385444	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	
50) Toluene-d8	10.09	98	1412585	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
62) 4-Bromofluorobenzene	12.40	95	380149	38.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.78%	
Target Compounds						
16) Acetone	3.82	43	50688	4.274	ug/l	98
18) Methyl Acetate	4.33	43	14066	2.261	ug/l #	73
20) Methylene Chloride	4.55	84	33466	4.140	ug/l	93
43) Isopropyl Acetate	8.17	43	262106	21.416	ug/l	95
95) Naphthalene	15.13	128	21587	9.580	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051421.D
Acq On : 23 Sep 2018 18:28
Operator : MD\SY
Sample : J5023-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NB-PIT-1

Quant Time: Sep 24 07:49:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

