

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051433.D
 Acq On : 24 Sep 2018 00:11
 Operator : MD\SY
 Sample : PB113124ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124ZHE#04

Quant Time: Sep 24 08:16:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	639068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1006742	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	838915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	314528	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	437405	54.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.28%	
35) Dibromofluoromethane	7.59	113	379994	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
50) Toluene-d8	10.09	98	1388862	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
62) 4-Bromofluorobenzene	12.40	95	380218	38.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.54%	
Target Compounds						
18) Methyl Acetate	4.32	43	12046	1.976	ug/l	94
20) Methylene Chloride	4.55	84	11210	1.415	ug/l	86
43) Isopropyl Acetate	8.17	43	284329	23.458	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051433.D
Acq On : 24 Sep 2018 00:11
Operator : MD\SY
Sample : PB113124ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113124ZHE#04

Quant Time: Sep 24 08:16:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

