

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051437.D
 Acq On : 24 Sep 2018 1:49
 Operator : MD\SY
 Sample : PB113124ZHE#08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124ZHE#08

Quant Time: Sep 24 08:30:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	660221	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1041381	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	883160	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	322873	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	455072	55.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.04%	
35) Dibromofluoromethane	7.59	113	395684	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
50) Toluene-d8	10.09	98	1437382	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
62) 4-Bromofluorobenzene	12.40	95	387821	38.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.46%	
Target Compounds						
18) Methyl Acetate	4.33	43	15265	2.423	ug/l	95
20) Methylene Chloride	4.55	84	9628	1.176	ug/l	90
43) Isopropyl Acetate	8.17	43	281608	22.460	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051437.D
Acq On : 24 Sep 2018 1:49
Operator : MD\SY
Sample : PB113124ZHE#08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113124ZHE#08

Quant Time: Sep 24 08:30:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

