

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051438.D
 Acq On : 24 Sep 2018 2:14
 Operator : MD\SY
 Sample : PB113124ZHE#09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124ZHE#09

Quant Time: Sep 24 08:32:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	592317	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	963969	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	798505	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	282975	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	421791	56.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.70%	
35) Dibromofluoromethane	7.59	113	369442	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
50) Toluene-d8	10.09	98	1319507	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
62) 4-Bromofluorobenzene	12.40	95	356464	37.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.92%	
Target Compounds						
18) Methyl Acetate	4.33	43	19456	3.443	ug/l #	68
20) Methylene Chloride	4.55	84	12808	1.744	ug/l	89
43) Isopropyl Acetate	8.17	43	291081	25.080	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051438.D
Acq On : 24 Sep 2018 2:14
Operator : MD\SY
Sample : PB113124ZHE#09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113124ZHE#09

Quant Time: Sep 24 08:32:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

