

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051439.D
 Acq On : 24 Sep 2018 2:38
 Operator : MD\SY
 Sample : PB113124ZHE#10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124ZHE#10

Quant Time: Sep 24 08:37:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	607851	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	959959	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	776158	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	261163	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	418838	55.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.02%	
35) Dibromofluoromethane	7.59	113	374600	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
50) Toluene-d8	10.09	98	1313838	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
62) 4-Bromofluorobenzene	12.40	95	344570	36.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.70%	
Target Compounds						
18) Methyl Acetate	4.34	43	13622	2.349	ug/l	99
20) Methylene Chloride	4.55	84	9362	1.242	ug/l #	87
43) Isopropyl Acetate	8.17	43	293422	25.388	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051439.D
Acq On : 24 Sep 2018 2:38
Operator : MD\SY
Sample : PB113124ZHE#10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113124ZHE#10

Quant Time: Sep 24 08:37:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

