

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
 Data File : VN051442.D
 Acq On : 24 Sep 2018 3:52
 Operator : MD\SY
 Sample : PB113124ZHE#13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113124ZHE#13

Quant Time: Sep 24 08:41:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	595489	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	948768	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	788423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	278113	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	414509	55.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.14%	
35) Dibromofluoromethane	7.59	113	370924	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
50) Toluene-d8	10.09	98	1309110	45.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.64%	
62) 4-Bromofluorobenzene	12.40	95	352554	38.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.30%	
Target Compounds						
18) Methyl Acetate	4.33	43	15097	2.657	ug/l	99
20) Methylene Chloride	4.56	84	10764	1.458	ug/l	84
43) Isopropyl Acetate	8.17	43	288066	25.218	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092318\
Data File : VN051442.D
Acq On : 24 Sep 2018 3:52
Operator : MD\SY
Sample : PB113124ZHE#13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB113124ZHE#13

Quant Time: Sep 24 08:41:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

