

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092318\
 Data File : VN051399.D
 Acq On : 22 Sep 2018 15:56
 Operator : MD\SY
 Sample : VSTDICV050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN092318

Quant Time: Sep 24 06:54:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.466	0.450	3.4	99	0.00
3 P	Chloromethane	0.743	0.734	1.2	101	0.00
4 C	Vinyl Chloride	0.820	0.808	1.5#	101	0.00
5 T	Bromomethane	0.524	0.507	3.2	101	0.00
6 T	Chloroethane	0.511	0.506	1.0	97	0.00
7 T	Trichlorofluoromethane	1.074	1.106	-3.0	99	0.00
8 T	Diethyl Ether	0.353	0.402	-13.9	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.606	0.543	10.4	76	0.00
10 T	Methyl Iodide	0.757	0.759	-0.3	79	0.00
11 T	Tert butyl alcohol	0.035	0.035	0.0	86	0.00
12 CM	1,1-Dichloroethene	0.556	0.508	8.6#	80	0.00
13 T	Acrolein	0.035	0.033	5.7	77	0.00
14 T	Allyl chloride	0.933	0.904	3.1	79	0.00
15 T	Acrylonitrile	0.192	0.187	2.6	80	0.00
16 T	Acetone	0.147	0.130	11.6	74	0.00
17 T	Carbon Disulfide	1.634	1.522	6.9	78	0.00
18 T	Methyl Acetate	0.477	0.457	4.2	79	0.00
19 T	Methyl tert-butyl Ether	1.435	1.442	-0.5	89	0.00
20 T	Methylene Chloride	0.620	0.583	6.0	78	0.00
21 T	trans-1,2-Dichloroethene	0.569	0.553	2.8	86	0.00
22 T	Diisopropyl ether	1.804	1.868	-3.5	100	0.00
23 T	Vinyl Acetate	1.184	1.223	-3.3	100	0.00
24 P	1,1-Dichloroethane	1.076	1.083	-0.7	99	0.00
25 T	2-Butanone	0.214	0.206	3.7	96	0.00
26 T	2,2-Dichloropropane	0.840	0.819	2.5	98	0.00
27 T	cis-1,2-Dichloroethene	0.624	0.618	1.0	96	0.00
28 T	Bromochloromethane	0.497	0.481	3.2	94	0.00
29 T	Tetrahydrofuran	0.136	0.142	-4.4	101	0.00
30 C	Chloroform	1.067	1.076	-0.8#	98	0.00
31 T	Cyclohexane	0.986	0.973	1.3	98	0.00
32 T	1,1,1-Trichloroethane	0.912	0.930	-2.0	100	0.00
33 S	1,2-Dichloroethane-d4	0.626	0.596	4.8	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.400	0.378	5.5	96	0.00
36 T	1,1-Dichloropropene	0.579	0.574	0.9	99	0.00
37 T	Ethyl Acetate	0.333	0.338	-1.5	100	0.00
38 T	Carbon Tetrachloride	0.606	0.587	3.1	95	0.00
39 T	Methylcyclohexane	0.677	0.671	0.9	97	0.00
40 TM	Benzene	1.700	1.701	-0.1	98	0.00
41 T	Methacrylonitrile	0.184	0.193	-4.9	108	0.02
42 TM	1,2-Dichloroethane	0.544	0.546	-0.4	101	0.00
43 T	Isopropyl Acetate	0.602	0.608	-1.0	102	0.00
44 TM	Trichloroethene	0.452	0.446	1.3	98	0.00
45 C	1,2-Dichloropropane	0.462	0.461	0.2#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092318\
 Data File : VN051399.D
 Acq On : 22 Sep 2018 15:56
 Operator : MD\SY
 Sample : VSTDICV050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN092318

Quant Time: Sep 24 06:54:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 24 06:41:32 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.261	0.262	-0.4	99	0.00
47 T	Bromodichloromethane	0.578	0.577	0.2	98	0.00
48 T	Methyl methacrylate	0.300	0.307	-2.3	102	0.00
49 T	1,4-Dioxane	0.004	0.004	0.0	100	0.00
50 S	Toluene-d8	1.506	1.401	7.0	94	0.00
51 T	4-Methyl-2-Pentanone	0.331	0.330	0.3	99	0.00
52 CM	Toluene	1.057	1.039	1.7#	98	0.00
53 T	t-1,3-Dichloropropene	0.583	0.579	0.7	99	0.00
54 T	cis-1,3-Dichloropropene	0.677	0.673	0.6	98	0.00
55 T	1,1,2-Trichloroethane	0.373	0.373	0.0	98	0.00
56 T	Ethyl methacrylate	0.457	0.472	-3.3	100	0.00
57 T	1,3-Dichloropropane	0.640	0.634	0.9	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.228	0.241	-5.7	102	0.00
59 T	2-Hexanone	0.215	0.211	1.9	97	0.00
60 T	Dibromochloromethane	0.441	0.435	1.4	97	0.00
61 T	1,2-Dibromoethane	0.362	0.360	0.6	100	0.00
62 S	4-Bromofluorobenzene	0.487	0.456	6.4	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.454	0.442	2.6	96	0.00
65 PM	Chlorobenzene	1.283	1.256	2.1	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.489	0.484	1.0	99	0.00
67 C	Ethyl Benzene	2.162	2.175	-0.6#	98	0.00
68 T	m/p-Xylenes	0.836	0.838	-0.2	96	0.00
69 T	o-Xylene	0.804	0.802	0.2	96	0.00
70 T	Styrene	1.274	1.296	-1.7	97	0.00
71 P	Bromoform	0.332	0.324	2.4	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	4.219	4.085	3.2	97	0.00
74 T	N-amyl acetate	1.058	1.042	1.5	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.984	0.925	6.0	98	0.00
76 T	1,2,3-Trichloropropane	0.830	0.826	0.5	99	0.00
77 T	Bromobenzene	1.068	1.007	5.7	98	0.00
78 T	n-propylbenzene	4.755	4.603	3.2	96	0.00
79 T	2-Chlorotoluene	2.846	2.717	4.5	98	0.00
80 T	1,3,5-Trimethylbenzene	3.440	3.362	2.3	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.270	0.254	5.9	96	0.00
82 T	4-Chlorotoluene	2.841	2.691	5.3	96	0.00
83 T	tert-Butylbenzene	3.062	3.042	0.7	100	0.00
84 T	1,2,4-Trimethylbenzene	3.466	3.536	-2.0	101	0.00
85 T	sec-Butylbenzene	4.142	4.204	-1.5	102	0.00
86 T	p-Isopropyltoluene	3.584	3.652	-1.9	101	0.00
87 T	1,3-Dichlorobenzene	1.848	1.809	2.1	100	0.00
88 T	1,4-Dichlorobenzene	1.778	1.750	1.6	99	0.00
89 T	n-Butylbenzene	2.947	2.984	-1.3	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092318\
Data File : VN051399.D
Acq On : 22 Sep 2018 15:56
Operator : MD\SY
Sample : VSTDICV050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN092318

Quant Time: Sep 24 06:54:31 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N092318W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 24 06:41:32 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.714	0.692	3.1	101	0.00
91 T	1,2-Dichlorobenzene	1.793	1.760	1.8	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.140	0.139	0.7	102	0.00
93 T	1,2,4-Trichlorobenzene	0.872	0.871	0.1	97	0.00
94 T	Hexachlorobutadiene	0.678	0.644	5.0	98	0.00
95 T	Naphthalene	1.627	1.616	0.7	98	0.00
96 T	1,2,3-Trichlorobenzene	0.861	0.851	1.2	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6