

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058286.D
 Acq On : 23 Sep 2019 12:57
 Operator : JC/SP
 Sample : K4975-17 25X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP06-20190918

Quant Time: Sep 25 11:07:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	483529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	791454	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	662183	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	256691	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	362329	53.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.60%	
35) Dibromofluoromethane	7.58	113	254201	52.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.66%	
50) Toluene-d8	10.09	98	1007214	53.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.46%	
62) 4-Bromofluorobenzene	12.41	95	312504	44.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.76%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058286.D
Acq On : 23 Sep 2019 12:57
Operator : JC/SP
Sample : K4975-17 25X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP06-20190918

Quant Time: Sep 25 11:07:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

