

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058288.D
 Acq On : 23 Sep 2019 13:40
 Operator : JC/SP
 Sample : K4939-07ME
 Misc : 5.85g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 986-B-3-(24)ME

Manual Integrations
 APPROVED

MMDadoda
 9/25/2019 12:47:41 AM

Quant Time: Sep 25 11:09:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	450662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	724522	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	586826	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	239323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	323726	51.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.14%	
35) Dibromofluoromethane	7.57	113	218144	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
50) Toluene-d8	10.08	98	899562	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
62) 4-Bromofluorobenzene	12.41	95	277194	43.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.96%	
Target Compounds						
64) Tetrachloroethene	10.63	164	12201	4.265	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058288.D
Acq On : 23 Sep 2019 13:40
Operator : JC/SP
Sample : K4939-07ME
Misc : 5.85g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
986-B-3-(24)ME

Manual Integrations
APPROVED

MMDadoda
9/25/2019 12:47:41 AM

Quant Time: Sep 25 11:09:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

