

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058293.D
 Acq On : 23 Sep 2019 15:29
 Operator : JC/SP
 Sample : K4939-01MEDL 5X
 Misc : 6.52g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 982-S-3-(19)MEDL

Quant Time: Sep 25 11:17:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	436826	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	706898	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	606090	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	245205	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	319749	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
35) Dibromofluoromethane	7.57	113	217946	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
50) Toluene-d8	10.08	98	904937	54.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.10%	
62) 4-Bromofluorobenzene	12.41	95	287217	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
Target Compounds						
64) Tetrachloroethene	10.63	164	114324	38.694	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058293.D
Acq On : 23 Sep 2019 15:29
Operator : JC/SP
Sample : K4939-01MEDL 5X
Misc : 6.52g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
982-S-3-(19)MEDL

Quant Time: Sep 25 11:17:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

