

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092319\
 Data File : VN058295.D
 Acq On : 23 Sep 2019 16:14
 Operator : JC/SP
 Sample : K4975-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE139D2-20190918

Quant Time: Oct 01 08:19:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	487347	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	793441	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	673802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	254795	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	360435	53.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.20%	
35) Dibromofluoromethane	7.58	113	250520	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
50) Toluene-d8	10.08	98	1011218	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
62) 4-Bromofluorobenzene	12.41	95	313404	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.73	101	6327	1.606	ug/l	94
12) 1,1-Dichloroethene	3.72	96	5370	1.450	ug/l #	75
27) cis-1,2-Dichloroethene	6.81	96	6736	1.363	ug/l	39
30) Chloroform	7.35	83	4008	0.426	ug/l	95
38) Carbon Tetrachloride	7.76	117	8728	1.517	ug/l	86
44) Trichloroethene	8.82	130	672966	157.355	ug/l	99
55) 1,1,2-Trichloroethane	10.56	97	3163	0.651	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092319\
Data File : VN058295.D
Acq On : 23 Sep 2019 16:14
Operator : JC/SP
Sample : K4975-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE139D2-20190918

Quant Time: Oct 01 08:19:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

