

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN092319\
 Data File : VN058302.D
 Acq On : 23 Sep 2019 18:50
 Operator : JC/SP
 Sample : K4975-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP05-20190917

Quant Time: Oct 01 08:43:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	479537	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	784735	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	668884	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	252890	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	360839	54.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.06%	
35) Dibromofluoromethane	7.58	113	249714	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
50) Toluene-d8	10.09	98	995290	53.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.10%	
62) 4-Bromofluorobenzene	12.41	95	310552	44.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.96%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.74	101	2410	0.622	ug/l #	84
16) Acetone	3.81	43	14605	6.442	ug/l	93
44) Trichloroethene	8.82	130	119240	28.190	ug/l	98
64) Tetrachloroethene	10.63	164	3072	0.942	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN092319\
Data File : VN058302.D
Acq On : 23 Sep 2019 18:50
Operator : JC/SP
Sample : K4975-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP05-20190917

Quant Time: Oct 01 08:43:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

