

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058314.D
 Acq On : 24 Sep 2019 00:01
 Operator : JC/SP
 Sample : K5002-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S49-0516

Manual Integrations
 APPROVED

MMDadoda
 9/25/2019 12:49:41 AM

Quant Time: Sep 24 06:54:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	462607	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	750306	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	647251	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	267310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	351805	54.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.20%	
35) Dibromofluoromethane	7.58	113	238466	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
50) Toluene-d8	10.08	98	966985	54.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.82%	
62) 4-Bromofluorobenzene	12.41	95	307128	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
Target Compounds						
16) Acetone	3.80	43	45284m	20.705	ug/l	Qvalue
25) 2-Butanone	6.83	43	14451m	4.850	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058314.D
Acq On : 24 Sep 2019 00:01
Operator : JC/SP
Sample : K5002-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S49-0516

Manual Integrations
APPROVED

MMDadoda
9/25/2019 12:49:41 AM

Quant Time: Sep 24 06:54:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

