

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058315.D
 Acq On : 24 Sep 2019 00:24
 Operator : JC/SP
 Sample : K5002-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S49-0521

Manual Integrations
 APPROVED

MMDadoda
 9/25/2019 12:49:43 AM

Quant Time: Sep 25 12:04:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	481023	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	792269	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	675025	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	270102	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	358409	53.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.00%	
35) Dibromofluoromethane	7.58	113	248951	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
50) Toluene-d8	10.09	98	1002881	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
62) 4-Bromofluorobenzene	12.41	95	310639	44.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.12%	
Target Compounds						
4) Vinyl Chloride	2.17	62	7207m	1.281	ug/l	
12) 1,1-Dichloroethene	3.72	96	3544m	0.851	ug/l	
24) 1,1-Dichloroethane	5.83	63	47170	5.468	ug/l	97
27) cis-1,2-Dichloroethene	6.82	96	142041	29.119	ug/l	89
44) Trichloroethene	8.83	130	3262	0.764	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058315.D
Acq On : 24 Sep 2019 00:24
Operator : JC/SP
Sample : K5002-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S49-0521

Manual Integrations
APPROVED

MMDadoda
9/25/2019 12:49:43 AM

Quant Time: Sep 25 12:04:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

