

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058316.D
 Acq On : 24 Sep 2019 00:46
 Operator : JC/SP
 Sample : K5002-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S49-0517

Manual Integrations
 APPROVED

MMDadoda
 9/25/2019 12:49:44 AM

Quant Time: Sep 24 09:12:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	460577	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	768132	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	656678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	257556	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	355860	55.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.94%	
35) Dibromofluoromethane	7.58	113	240928	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
50) Toluene-d8	10.08	98	966643	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
62) 4-Bromofluorobenzene	12.41	95	304821	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
Target Compounds						
21) trans-1,2-Dichloroethene	5.03	96	11265m	3.282	ug/l	
27) cis-1,2-Dichloroethene	6.82	96	50967	10.912	ug/l	87
44) Trichloroethene	8.83	130	13144	3.175	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058316.D
Acq On : 24 Sep 2019 00:46
Operator : JC/SP
Sample : K5002-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S49-0517

Manual Integrations
APPROVED

MMDadoda
9/25/2019 12:49:44 AM

Quant Time: Sep 24 09:12:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

