

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058317.D
 Acq On : 24 Sep 2019 1:08
 Operator : JC/SP
 Sample : K5002-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S49-0519

Quant Time: Sep 25 12:07:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	454633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	742555	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	623463	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	233476	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	340366	53.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.50%	
35) Dibromofluoromethane	7.58	113	236641	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
50) Toluene-d8	10.08	98	942840	53.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.22%	
62) 4-Bromofluorobenzene	12.41	95	287872	44.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.12%	
Target Compounds						
21) trans-1,2-Dichloroethene	5.02	96	2636	0.468	ug/l #	79
27) cis-1,2-Dichloroethene	6.81	96	22427	4.864	ug/l	88
44) Trichloroethene	8.82	130	19447	4.859	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058317.D
Acq On : 24 Sep 2019 1:08
Operator : JC/SP
Sample : K5002-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S49-0519

Quant Time: Sep 25 12:07:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

