

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058321.D
 Acq On : 24 Sep 2019 2:37
 Operator : JC/SP
 Sample : K4997-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-09-091919

Manual Integrations
 APPROVED

MMDadoda
 9/25/2019 12:49:48 AM

Quant Time: Sep 24 09:15:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	469594	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	793155	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	703299	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	287697	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	358737	54.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.70%	
35) Dibromofluoromethane	7.58	113	250285	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
50) Toluene-d8	10.09	98	1016212	54.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.18%	
62) 4-Bromofluorobenzene	12.41	95	333908	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
Target Compounds						
16) Acetone	3.80	43	13147m	5.922	ug/l	
20) Methylene Chloride	4.53	84	59807	15.232	ug/l	97
43) Isopropyl Acetate	8.15	43	97186	8.767	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058321.D
Acq On : 24 Sep 2019 2:37
Operator : JC/SP
Sample : K4997-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-09-091919

Manual Integrations
APPROVED

MMDadoda
9/25/2019 12:49:48 AM

Quant Time: Sep 24 09:15:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

