

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058323.D
 Acq On : 24 Sep 2019 3:21
 Operator : JC/SP
 Sample : K4997-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-11-091919

Quant Time: Sep 24 09:16:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	470755	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	795909	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	701159	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	288652	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	364317	55.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.12%	
35) Dibromofluoromethane	7.58	113	247493	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
50) Toluene-d8	10.09	98	1019322	54.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.14%	
62) 4-Bromofluorobenzene	12.41	95	333938	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
Target Compounds						
16) Acetone	3.80	43	12118	5.445	ug/l	92
20) Methylene Chloride	4.53	84	56960	14.460	ug/l	98
43) Isopropyl Acetate	8.16	43	130019	11.688	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058323.D
Acq On : 24 Sep 2019 3:21
Operator : JC/SP
Sample : K4997-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-11-091919

Quant Time: Sep 24 09:16:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

