

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092319\
 Data File : VN058307.D
 Acq On : 23 Sep 2019 21:03
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 05:55:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	69	0.00
2 T	Dichlorodifluoromethane	0.411	0.351	14.6	64	0.00
3 P	Chloromethane	0.383	0.311	18.8	70	0.00
4 C	Vinyl Chloride	0.398	0.346	13.1#	70	0.00
5 T	Bromomethane	0.197	0.194	1.5	76	0.00
6 T	Chloroethane	0.263	0.237	9.9	73	0.00
7 T	Trichlorofluoromethane	0.561	0.521	7.1	73	0.00
8 T	Diethyl Ether	0.257	0.243	5.4	70	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.404	0.377	6.7	71	0.00
10 T	Methyl Iodide	0.339	0.344	-1.5	73	0.00
11 T	Tert butyl alcohol	0.083	0.097	-16.9	82	0.00
12 CM	1,1-Dichloroethene	0.342	0.304	11.1#	71	0.00
13 T	Acrolein	0.012	0.021	-75.0#	120	0.00
14 T	Allyl chloride	0.661	0.555	16.0	62	0.00
15 T	Acrylonitrile	0.221	0.260	-17.6	83	0.00
16 T	Acetone	0.236	0.226	4.2	72	0.00
17 T	Carbon Disulfide	0.526	0.363	31.0#	63	0.00
18 T	Methyl Acetate	0.622	0.694	-11.6	84	0.00
19 T	Methyl tert-butyl Ether	1.525	1.611	-5.6	76	0.00
20 T	Methylene Chloride	0.460	0.435	5.4	76	0.00
21 T	trans-1,2-Dichloroethene	0.371	0.326	12.1	70	0.00
22 T	Diisopropyl ether	1.680	1.776	-5.7	76	0.00
23 T	Vinyl Acetate	1.150	1.304	-13.4	76	0.00
24 P	1,1-Dichloroethane	0.897	0.913	-1.8	76	0.00
25 T	2-Butanone	0.322	0.360	-11.8	78	0.00
26 T	2,2-Dichloropropane	0.761	0.653	14.2	64	0.00
27 T	cis-1,2-Dichloroethene	0.507	0.501	1.2	72	0.00
28 T	Bromochloromethane	0.424	0.446	-5.2	71	0.00
29 T	Tetrahydrofuran	0.198	0.226	-14.1	83	0.00
30 C	Chloroform	0.965	0.997	-3.3#	77	0.00
31 T	Cyclohexane	0.664	0.558	16.0	68	0.00
32 T	1,1,1-Trichloroethane	0.735	0.759	-3.3	73	0.00
33 S	1,2-Dichloroethane-d4	0.696	0.730	-4.9	72	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	70	0.00
35 S	Dibromofluoromethane	0.304	0.311	-2.3	70	0.00
36 T	1,1-Dichloropropene	0.361	0.328	9.1	70	0.00
37 T	Ethyl Acetate	0.374	0.433	-15.8	82	0.00
38 T	Carbon Tetrachloride	0.362	0.355	1.9	72	0.00
39 T	Methylcyclohexane	0.380	0.322	15.3	67	0.00
40 TM	Benzene	1.101	1.092	0.8	74	0.00
41 T	Methacrylonitrile	0.181	0.187	-3.3	70	0.00
42 TM	1,2-Dichloroethane	0.462	0.479	-3.7	77	0.00
43 T	Isopropyl Acetate	0.699	0.743	-6.3	78	0.00
44 TM	Trichloroethene	0.270	0.265	1.9	72	0.00
45 C	1,2-Dichloropropane	0.329	0.346	-5.2#	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092319\
 Data File : VN058307.D
 Acq On : 23 Sep 2019 21:03
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 05:55:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.200	0.207	-3.5	75	0.00
47 T	Bromodichloromethane	0.419	0.451	-7.6	75	0.00
48 T	Methyl methacrylate	0.317	0.359	-13.2	79	0.00
49 T	1,4-Dioxane	0.005	0.006	-20.0	83	0.00
50 S	Toluene-d8	1.184	1.209	-2.1	68	0.00
51 T	4-Methyl-2-Pentanone	0.371	0.480	-29.4#	84	0.00
52 CM	Toluene	0.690	0.692	-0.3#	73	0.00
53 T	t-1,3-Dichloropropene	0.429	0.469	-9.3	73	0.00
54 T	cis-1,3-Dichloropropene	0.474	0.490	-3.4	71	0.00
55 T	1,1,2-Trichloroethane	0.306	0.335	-9.5	77	0.00
56 T	Ethyl methacrylate	0.442	0.500	-13.1	76	0.00
57 T	1,3-Dichloropropane	0.527	0.571	-8.3	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.211	0.243	-15.2	73	0.00
59 T	2-Hexanone	0.276	0.346	-25.4#	82	0.00
60 T	Dibromochloromethane	0.295	0.331	-12.2	74	0.00
61 T	1,2-Dibromoethane	0.284	0.312	-9.9	78	0.00
62 S	4-Bromofluorobenzene	0.440	0.448	-1.8	69	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	70	0.00
64 T	Tetrachloroethene	0.244	0.233	4.5	73	0.00
65 PM	Chlorobenzene	0.879	0.884	-0.6	74	0.00
66 T	1,1,1,2-Tetrachloroethane	0.319	0.342	-7.2	74	0.00
67 C	Ethyl Benzene	1.507	1.578	-4.7#	73	0.00
68 T	m/p-Xylenes	0.568	0.573	-0.9	72	0.00
69 T	o-Xylene	0.558	0.579	-3.8	74	0.00
70 T	Styrene	0.962	1.037	-7.8	74	0.00
71 P	Bromoform	0.203	0.235	-15.8	76	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
73 T	Isopropylbenzene	3.130	3.357	-7.3	74	0.00
74 T	N-amyl acetate	1.309	1.545	-18.0	78	0.00
75 P	1,1,2,2-Tetrachloroethane	1.048	1.193	-13.8	80	0.00
76 T	1,2,3-Trichloropropane	0.895	0.976	-9.1	79	0.00
77 T	Bromobenzene	0.787	0.798	-1.4	74	0.00
78 T	n-propylbenzene	3.556	3.942	-10.9	74	0.00
79 T	2-Chlorotoluene	2.223	2.357	-6.0	74	0.00
80 T	1,3,5-Trimethylbenzene	2.647	2.872	-8.5	73	0.00
81 T	trans-1,4-Dichloro-2-butene	0.271	0.287	-5.9	71	0.00
82 T	4-Chlorotoluene	2.336	2.498	-6.9	74	0.00
83 T	tert-Butylbenzene	2.360	2.556	-8.3	74	0.00
84 T	1,2,4-Trimethylbenzene	2.634	2.919	-10.8	74	0.00
85 T	sec-Butylbenzene	3.086	3.439	-11.4	74	0.00
86 T	p-Isopropyltoluene	2.744	3.095	-12.8	74	0.00
87 T	1,3-Dichlorobenzene	1.438	1.528	-6.3	74	0.00
88 T	1,4-Dichlorobenzene	1.468	1.521	-3.6	73	0.00
89 T	n-Butylbenzene	2.601	2.762	-6.2	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092319\
Data File : VN058307.D
Acq On : 23 Sep 2019 21:03
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 24 05:55:31 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.426	0.491	-15.3	75	0.00
91 T	1,2-Dichlorobenzene	1.442	1.563	-8.4	75	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.189	0.220	-16.4	79	0.00
93 T	1,2,4-Trichlorobenzene	0.922	0.917	0.5	67	0.00
94 T	Hexachlorobutadiene	0.445	0.433	2.7	68	0.00
95 T	Naphthalene	2.388	2.669	-11.8	71	0.00
96 T	1,2,3-Trichlorobenzene	0.871	0.885	-1.6	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6