

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092319\
 Data File : VN058334.D
 Acq On : 24 Sep 2019 8:30
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 09:31:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	0.411	0.356	13.4	68	0.00
3 P	Chloromethane	0.383	0.304	20.6	72	0.00
4 C	Vinyl Chloride	0.398	0.336	15.6#	71	0.00
5 T	Bromomethane	0.197	0.149	24.4	61	0.00
6 T	Chloroethane	0.263	0.225	14.4	73	0.00
7 T	Trichlorofluoromethane	0.561	0.501	10.7	73	0.00
8 T	Diethyl Ether	0.257	0.261	-1.6	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.404	0.369	8.7	73	0.00
10 T	Methyl Iodide	0.339	0.241	28.9#	54	0.00
11 T	Tert butyl alcohol	0.083	0.094	-13.3	84	0.00
12 CM	1,1-Dichloroethene	0.342	0.305	10.8#	74	0.00
13 T	Acrolein	0.012	0.015	-25.0	86	0.00
14 T	Allyl chloride	0.661	0.623	5.7	73	0.00
15 T	Acrylonitrile	0.221	0.254	-14.9	84	0.00
16 T	Acetone	0.236	0.217	8.1	73	0.00
17 T	Carbon Disulfide	0.526	0.384	27.0#	70	0.00
18 T	Methyl Acetate	0.622	0.707	-13.7	89	0.00
19 T	Methyl tert-butyl Ether	1.525	1.651	-8.3	81	0.00
20 T	Methylene Chloride	0.460	0.439	4.6	80	0.00
21 T	trans-1,2-Dichloroethene	0.371	0.332	10.5	74	0.00
22 T	Diisopropyl ether	1.680	1.801	-7.2	80	0.00
23 T	Vinyl Acetate	1.150	1.295	-12.6	79	0.00
24 P	1,1-Dichloroethane	0.897	0.920	-2.6	80	0.00
25 T	2-Butanone	0.322	0.352	-9.3	80	0.00
26 T	2,2-Dichloropropane	0.761	0.419	44.9#	43#	0.00
27 T	cis-1,2-Dichloroethene	0.507	0.518	-2.2	78	0.00
28 T	Bromochloromethane	0.424	0.471	-11.1	79	0.00
29 T	Tetrahydrofuran	0.198	0.219	-10.6	83	0.00
30 C	Chloroform	0.965	0.993	-2.9#	80	0.00
31 T	Cyclohexane	0.664	0.552	16.9	71	0.00
32 T	1,1,1-Trichloroethane	0.735	0.769	-4.6	77	0.00
33 S	1,2-Dichloroethane-d4	0.696	0.735	-5.6	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.304	0.303	0.3	75	0.00
36 T	1,1-Dichloropropene	0.361	0.317	12.2	75	0.00
37 T	Ethyl Acetate	0.374	0.398	-6.4	83	0.00
38 T	Carbon Tetrachloride	0.362	0.343	5.2	77	0.00
39 T	Methylcyclohexane	0.380	0.298	21.6	68	0.00
40 TM	Benzene	1.101	1.038	5.7	77	0.00
41 T	Methacrylonitrile	0.181	0.222	-22.7	91	0.00
42 TM	1,2-Dichloroethane	0.462	0.453	1.9	80	0.00
43 T	Isopropyl Acetate	0.699	0.741	-6.0	86	0.00
44 TM	Trichloroethene	0.270	0.250	7.4	75	0.00
45 C	1,2-Dichloropropane	0.329	0.328	0.3#	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092319\
 Data File : VN058334.D
 Acq On : 24 Sep 2019 8:30
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 09:31:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.200	0.199	0.5	79	0.00
47 T	Bromodichloromethane	0.419	0.438	-4.5	80	0.00
48 T	Methyl methacrylate	0.317	0.342	-7.9	83	0.00
49 T	1,4-Dioxane	0.005	0.006	-20.0	79	0.00
50 S	Toluene-d8	1.184	1.163	1.8	72	0.00
51 T	4-Methyl-2-Pentanone	0.371	0.445	-19.9	85	0.00
52 CM	Toluene	0.690	0.655	5.1#	76	0.00
53 T	t-1,3-Dichloropropene	0.429	0.427	0.5	73	0.00
54 T	cis-1,3-Dichloropropene	0.474	0.448	5.5	71	0.00
55 T	1,1,2-Trichloroethane	0.306	0.318	-3.9	81	0.00
56 T	Ethyl methacrylate	0.442	0.486	-10.0	81	0.00
57 T	1,3-Dichloropropane	0.527	0.531	-0.8	79	0.00
58 T	2-Chloroethyl Vinyl ether	0.211	0.242	-14.7	80	0.00
59 T	2-Hexanone	0.276	0.320	-15.9	83	0.00
60 T	Dibromochloromethane	0.295	0.320	-8.5	79	0.00
61 T	1,2-Dibromoethane	0.284	0.294	-3.5	80	0.00
62 S	4-Bromofluorobenzene	0.440	0.429	2.5	73	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	75	0.00
64 T	Tetrachloroethene	0.244	0.239	2.0	80	0.00
65 PM	Chlorobenzene	0.879	0.862	1.9	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.319	0.344	-7.8	80	0.00
67 C	Ethyl Benzene	1.507	1.519	-0.8#	76	0.00
68 T	m/p-Xylenes	0.568	0.551	3.0	75	0.00
69 T	o-Xylene	0.558	0.561	-0.5	77	0.00
70 T	Styrene	0.962	1.004	-4.4	76	0.00
71 P	Bromoform	0.203	0.234	-15.3	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.00
73 T	Isopropylbenzene	3.130	3.304	-5.6	75	0.00
74 T	N-amyl acetate	1.309	1.561	-19.3	81	0.00
75 P	1,1,2,2-Tetrachloroethane	1.048	1.186	-13.2	82	0.00
76 T	1,2,3-Trichloropropane	0.895	0.999	-11.6	83	0.00
77 T	Bromobenzene	0.787	0.791	-0.5	76	0.00
78 T	n-propylbenzene	3.556	3.820	-7.4	74	0.00
79 T	2-Chlorotoluene	2.223	2.340	-5.3	76	0.00
80 T	1,3,5-Trimethylbenzene	2.647	2.837	-7.2	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.271	0.255	5.9	65	0.00
82 T	4-Chlorotoluene	2.336	2.441	-4.5	74	0.00
83 T	tert-Butylbenzene	2.360	2.542	-7.7	76	0.00
84 T	1,2,4-Trimethylbenzene	2.634	2.893	-9.8	76	0.00
85 T	sec-Butylbenzene	3.086	3.337	-8.1	74	0.00
86 T	p-Isopropyltoluene	2.744	2.944	-7.3	72	0.00
87 T	1,3-Dichlorobenzene	1.438	1.496	-4.0	75	0.00
88 T	1,4-Dichlorobenzene	1.468	1.500	-2.2	74	0.00
89 T	n-Butylbenzene	2.601	2.648	-1.8	68	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN092319\
Data File : VN058334.D
Acq On : 24 Sep 2019 8:30
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 66 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 24 09:31:06 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.426	0.477	-12.0	75	0.00
91 T	1,2-Dichlorobenzene	1.442	1.532	-6.2	75	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.189	0.228	-20.6	84	0.00
93 T	1,2,4-Trichlorobenzene	0.922	0.978	-6.1	74	0.00
94 T	Hexachlorobutadiene	0.445	0.436	2.0	71	0.00
95 T	Naphthalene	2.388	2.940	-23.1	80	0.00
96 T	1,2,3-Trichlorobenzene	0.871	0.934	-7.2	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6