

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092319\
 Data File : VN058328.D
 Acq On : 24 Sep 2019 5:34
 Operator : JC/SP
 Sample : K4975-14DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE139D2-20190918DL

Quant Time: Sep 24 09:19:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 19 09:27:53 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	473494	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	791299	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	673903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	274931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	360183	54.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.24%	
35) Dibromofluoromethane	7.58	113	245717	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
50) Toluene-d8	10.08	98	1005671	53.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.32%	
62) 4-Bromofluorobenzene	12.41	95	324333	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
Target Compounds						
44) Trichloroethene	8.82	130	120903	28.346	ug/l	Qvalue 98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092319\
Data File : VN058328.D
Acq On : 24 Sep 2019 5:34
Operator : JC/SP
Sample : K4975-14DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE139D2-20190918DL

Quant Time: Sep 24 09:19:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091819W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 19 09:27:53 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.65 min Scan# 1828

Delta R.T. -0.00 min

Lab File: VN058328.D

Acq: 24 Sep 2019 5:34

Instrument :

MSVOA_N

ClientSampleId :


RE139D2-20190918DL

Tot Ion:168 Resp: 473494

Ion Ratio Lower Upper

168 100

99 62.2 47.4 71.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.65

200000

150000



100000

50000

0

7.60 7.70

Time-->



#33

1,2-Dichloroethane-d4

Concen: 54.616 ug/l

RT: 8.01 min Scan# 1941

Delta R.T. 0.00 min

Lab File: VN058328.D


Acq: 24 Sep 2019 5:34

Tgt Ion: 65 Resp: 360183

Ion Ratio Lower Upper

65 100

67 51.5 0.0 103.4



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO

8.01

150000


100000


50000

0

7.90 8.00 8.10

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2115
 Delta R.T. -0.00 min
 Lab File: VN058328.D
 Acq: 24 Sep 2019 5:34

Instrument :
 MSVOA_N
 ClientSampleId :
 RE139D2-20190918DL

Tot Ion	Ratio	Lower	Upper
114	100		
63	22.1	0.0	44.2
88	15.5	0.0	31.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 51.080 ug/l
 RT: 7.58 min Scan# 1805
 Delta R.T. -0.00 min
 Lab File: VN058328.D
 Acq: 24 Sep 2019 5:34

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.6	81.8	122.6
192	17.7	14.5	21.7



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#44

Trichloroethene

Concen: 28.346 ug/l

RT: 8.82 min Scan# 2193

Delta R.T. -0.00 min

Lab File: VN058328.D

Acq: 24 Sep 2019 5:34

Instrument :

MSVOA_N

ClientSampleId :


RE139D2-20190918DL

Tot Ion:130 Resp: 120903

Ion Ratio Lower Upper




130 100

95 101.9 0.0 207.8



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VN



#50

Toluene-d8

Concen: 53.657 ug/l

RT: 10.08 min Scan# 2585

Delta R.T. -0.00 min

Lab File: VN058328.D


Acq: 24 Sep 2019 5:34

Tgt Ion: 98 Resp: 1005671

Ion Ratio Lower Upper



98 100


100 63.0 51.1 76.7



Abundance Ion 98.00 (97.70 to 98.70): VN

Ion 100.00 (99.70 to 100.70): VN





#62

4-Bromofluorobenzene

Concen: 46.585 ug/l

RT: 12.41 min Scan# 3307

Delta R.T. 0.00 min

Lab File: VN058328.D

Acq: 24 Sep 2019 5:34

Instrument :

MSVOA_N

ClientSampleId :

RE139D2-20190918DL





Tot Ion: 95 Resp: 324333

Ion Ratio Lower Upper

95 100

174 73.8 0.0 152.2

176 71.6 0.0 148.0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. 0.00 min

Lab File: VN058328.D

Acq: 24 Sep 2019 5:34




Tgt Ion: 117 Resp: 673903




Ion Ratio Lower Upper

117 100

82 59.4 46.9 70.3

119 32.2 25.3 37.9





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3601

Delta R.T. 0.00 min

Lab File: VN058328.D

Acq: 24 Sep 2019 5:34

Instrument :

MSVOA_N

ClientSampleId :

RE139D2-20190918DL

Tot Ion:152 Resp: 274931

Ion Ratio Lower Upper

152 100

115 61.5 30.1 90.3

150 157.9 0.0 346.4

