

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063587.D
 Acq On : 23 Sep 2020 14:31
 Operator : JC/MD
 Sample : L4038-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: Sep 24 03:25:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	211203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	397912	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	402683	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	165614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	180038	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
35) Dibromofluoromethane	7.55	113	133013	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
50) Toluene-d8	10.06	98	515354	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.38	95	194866	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
Target Compounds						
16) Acetone	3.78	43	11735	10.156	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063587.D
Acq On : 23 Sep 2020 14:31
Operator : JC/MD
Sample : L4038-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: Sep 24 03:25:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

