

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN092320\
 Data File : VN063593.D
 Acq On : 23 Sep 2020 16:54
 Operator : JC/MD
 Sample : L4116-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KY038PS038-200922

Quant Time: Sep 24 03:46:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	207637	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	389089	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	399423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	166529	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	178176	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
35) Dibromofluoromethane	7.55	113	130052	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
50) Toluene-d8	10.06	98	508372	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.38	95	197974	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.78	96	22396	7.622	ug/l	86
44) Trichloroethene	8.80	130	1665	0.580	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN092320\
Data File : VN063593.D
Acq On : 23 Sep 2020 16:54
Operator : JC/MD
Sample : L4116-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KY038PS038-200922

Quant Time: Sep 24 03:46:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

